
Midmark M11 - v1.0.0
Total Cycles: 5406

Name: RITTER 1
Sterilizer ID: V2178198
03/11/2025 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 77.5 F 0.4
4:00 86.4 F 1.8
6:00 122.0 F 4.4
8:00 154.7 F 7.2
10:00 179.4 F 10.5
12:00 190.4 F 7.3
14:00 204.2 F 4.9
16:00 216.4 F 8.4
18:00 226.9 F 12.3
20:00 229.2 F 7.5
22:00 241.3 F 12.6
24:00 239.2 F 10.3
26:00 252.0 F 16.6
28:00 262.8 F 22.8
30:00 272.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.8 F 29.3
2:00 271.6 F 28.4
3:00 272.8 F 29.4
4:00 272.1 F 28.7

Min 271.5 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 180.7 F 0.0
10:00 195.6 F 0.0
15:00 200.4 F 0.0
20:00 204.5 F 0.0
25:00 214.7 F 0.0
30:00 209.8 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 30:23
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:08:38
CYCLE COMPLETE

03/11/2025 10:04 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____