
Midmark M11 - v1.0.0
Total Cycles: 3855

Name: RITTER 1
Sterilizer ID: V2178198
10/03/2023 02:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.1 F 0.0
2:00 87.6 F 0.0
4:00 100.6 F 0.0
6:00 117.1 F 0.0
8:00 165.0 F 0.0
10:00 184.5 F 2.9
12:00 201.3 F 6.6
14:00 215.0 F 10.2
16:00 223.2 F 8.8
18:00 233.1 F 8.5
20:00 243.4 F 13.2
22:00 240.3 F 10.5
24:00 235.5 F 8.4
26:00 250.9 F 15.4
28:00 262.5 F 22.1
30:00 272.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.0 F 28.1
2:00 272.4 F 28.5
3:00 272.4 F 28.9
4:00 271.4 F 28.1

Min 271.3 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.1 F 0.0
5:00 241.6 F 0.0
10:00 220.1 F 0.0
15:00 232.0 F 0.0
20:00 230.4 F 0.0
25:00 232.0 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:14
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:08:20
CYCLE COMPLETE

10/03/2023 03:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____