
Midmark M11 - v1.0.5
Total Cycles: 1785

Name: RITTER2
Sterilizer ID: V2313228
11/29/2022 01:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.0 F 0.0
2:00 162.8 F 0.0
4:00 157.3 F 0.4
6:00 186.4 F 3.4
8:00 206.7 F 7.5
10:00 235.0 F 12.8
12:00 234.5 F 9.1
14:00 239.5 F 10.2
16:00 245.1 F 13.1
18:00 244.6 F 12.5
20:00 257.5 F 19.6
22:00 269.2 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 272.0 F 29.0
2:00 272.4 F 29.1
3:00 272.9 F 29.4
4:00 272.6 F 29.6

Min 271.8 F 29.0
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 230.6 F 0.0
10:00 225.8 F 0.0
15:00 212.6 F 0.0
20:00 194.3 F 0.0
25:00 214.1 F 0.0
30:00 215.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 23:38
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:48
CYCLE COMPLETE

11/29/2022 02:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____