
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 007653
13:14 10/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:05
STERILIZING 3:05
136.0 C 220kPa 3:05
136.4 C 224kPa 3:35
136.4 C 224kPa 4:05
136.3 C 223kPa 4:35
136.3 C 223kPa 5:05
136.0 C 222kPa 5:35
136.3 C 223kPa 6:05
136.4 C 219kPa 6:35
136.4 C 219kPa 6:35
VENTING 6:35
AIR DRYING 7:27
CYCLE COMPLETE 37:27

Digital Signature #
2F99E7B5BF9E3DB2
