
Midmark M11 - v1.0.0
Total Cycles: 4135

Name: RITTER 1
Sterilizer ID: V2178198
12/19/2023 11:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.0 F 0.0
2:00 162.7 F 2.0
4:00 160.0 F 5.2
6:00 178.7 F 8.6
8:00 197.6 F 12.7
10:00 211.7 F 5.7
12:00 222.6 F 9.9
14:00 230.9 F 9.2
16:00 236.9 F 10.7
18:00 234.6 F 8.1
20:00 247.3 F 14.2
22:00 258.3 F 20.2
24:00 268.4 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.5 F 29.5
2:00 272.0 F 29.0
3:00 271.1 F 28.7
4:00 272.2 F 29.5

Min 271.1 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 218.4 F 0.0
10:00 225.3 F 0.0
15:00 229.1 F 0.0
20:00 229.9 F 0.0
25:00 229.2 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 25:18
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:04:07
CYCLE COMPLETE

12/19/2023 12:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____