
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001608
15:02 11/22/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:25
PRESSURIZING 6:04
STERILIZING 7:32
136.0 C 220kPa 7:32
136.4 C 223kPa 8:02
136.5 C 220kPa 8:32
136.4 C 223kPa 9:02
136.3 C 222kPa 9:32
135.9 C 217kPa 10:02
136.2 C 221kPa 10:32
136.4 C 222kPa 11:02
VENTING 11:02
AIR DRYING 12:04
CYCLE COMPLETE 42:04

Digital Signature #
CEEA5095C8139300
