

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2250

Name: RITTER 1  
Sterilizer ID: V2178198  
08/25/2022 02:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.2 F 0.0  
2:00 194.2 F 0.5  
4:00 208.9 F 2.1  
6:00 224.2 F 7.3  
8:00 238.0 F 12.3  
10:00 236.3 F 9.5  
12:00 238.7 F 9.1  
14:00 243.2 F 11.2  
16:00 247.9 F 13.8  
18:00 260.4 F 20.8  
20:00 271.8 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.4 F 28.4  
2:00 272.0 F 28.4  
3:00 272.1 F 28.2  
4:00 272.2 F 28.6

Min 271.4 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 218.7 F 0.0  
10:00 234.0 F 0.0  
15:00 235.3 F 0.0  
20:00 227.1 F 0.0  
25:00 229.9 F 0.0  
30:00 225.0 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 20:34  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:18  
CYCLE COMPLETE

08/25/2022 03:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_