
Midmark M11 - v1.0.0
Total Cycles: 2586

Name: RITTER 1
Sterilizer ID: V2178198
11/21/2022 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.3 F 0.0
2:00 197.2 F 4.9
4:00 204.0 F 10.3
6:00 220.1 F 5.3
8:00 235.4 F 11.3
10:00 234.7 F 8.9
12:00 239.7 F 9.5
14:00 248.5 F 14.7
16:00 260.9 F 21.5
18:00 272.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.6 F 29.1
2:00 271.7 F 28.4
3:00 272.3 F 29.3
4:00 272.4 F 28.9

Min 271.4 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 240.3 F 0.0
10:00 233.4 F 0.0
15:00 228.7 F 0.0
20:00 227.9 F 0.0
25:00 231.0 F 0.0
30:00 229.6 F 0.0

Drying Complete

FILLING: 2:13
HEATING: 18:33
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:01
CYCLE COMPLETE

11/21/2022 02:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____