
Midmark M11 - v1.0.5
Total Cycles: 1545

Name: RITTER2
Sterilizer ID: V2313228
09/27/2022 01:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.5 F 0.0
2:00 187.9 F 3.5
4:00 197.7 F 9.0
6:00 216.1 F 7.3
8:00 230.8 F 10.6
10:00 233.4 F 8.7
12:00 239.1 F 10.1
14:00 248.4 F 15.3
16:00 261.3 F 22.6
18:00 272.8 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.8 F 29.9
2:00 271.7 F 29.4
3:00 272.9 F 30.0
4:00 272.0 F 29.3

Min 271.7 F 29.2
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 197.1 F 0.0
10:00 209.4 F 0.0
15:00 215.8 F 0.0
20:00 219.3 F 0.0
25:00 220.4 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 3:57
HEATING: 18:21
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:32
CYCLE COMPLETE

09/27/2022 02:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____