

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001378  
10:31 07/12/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:53  
PRESSURIZING 3:56  
STERILIZING 5:13  
136.5 C 224kPa 5:13  
136.3 C 223kPa 5:43  
136.2 C 223kPa 6:13  
136.2 C 223kPa 6:43  
136.4 C 223kPa 7:13  
136.3 C 222kPa 7:43  
136.4 C 225kPa 8:13  
136.4 C 225kPa 8:43  
136.4 C 225kPa 8:43  
VENTING 8:43  
AIR DRYING 9:36  
CYCLE COMPLETE 18:22  
-----

Digital Signature #  
9BCE3B3BE1AC5B90  
-----