
Midmark M11 - v1.0.5
Total Cycles: 1069

Name: RITTER2
Sterilizer ID: V2313228
05/04/2022 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.2 F 0.0
2:00 171.4 F 2.7
4:00 182.1 F 7.1
6:00 206.2 F 12.3
8:00 219.5 F 7.4
10:00 236.6 F 12.8
12:00 237.1 F 10.5
14:00 236.2 F 8.9
16:00 250.1 F 15.9
18:00 262.2 F 22.8
20:00 272.7 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.4 F 29.7
2:00 271.4 F 28.9
3:00 272.0 F 29.9
4:00 272.2 F 29.4

Min 271.2 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 216.9 F 0.0
10:00 195.0 F 0.0
15:00 193.8 F 0.0
20:00 195.3 F 0.0
25:00 209.3 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 20:21
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

05/04/2022 09:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____