
Midmark M11 - v1.0.5
Total Cycles: 289

Name: RITTER2
Sterilizer ID: V2313228
08/19/2021 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.0 F 0.0
2:00 192.9 F 4.6
4:00 215.7 F 10.9
6:00 229.7 F 7.8
8:00 240.1 F 11.3
10:00 245.2 F 12.9
12:00 247.0 F 13.9
14:00 260.7 F 21.6
16:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.1 F 28.8
2:00 272.5 F 29.2
3:00 272.1 F 29.4
4:00 271.7 F 28.6

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 241.0 F 0.0
10:00 205.5 F 0.0
15:00 202.9 F 0.0
20:00 195.9 F 0.0
Dry Cycle Aborted

FILLING: 3:20
HEATING: 16:18
STERILIZING: 4:00
VENTING: 3:06
DRYING: 20:05
TOTAL CYCLE: 00:46:49

08/19/2021 10:43 AM
ACCEPTED: X
REJECTED:
Operator: U02