

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002516  
11:10 19/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:45  
PRESSURIZING 2:54  
STERILIZING 4:07  
136.5 C 220kPa 4:07  
136.3 C 224kPa 4:37  
136.2 C 222kPa 5:07  
136.4 C 223kPa 5:37  
136.1 C 223kPa 6:07  
136.3 C 223kPa 6:37  
136.4 C 223kPa 7:07  
136.4 C 224kPa 7:37  
136.4 C 224kPa 7:37  
VENTING 7:37  
AIR DRYING 8:29  
CYCLE COMPLETE 26:40  
-----

Digital Signature #  
4F269F13D8474511  
-----