

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2905

Name: RITTER 1  
Sterilizer ID: V2178198  
02/15/2023 09:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.9 F 0.0  
2:00 188.6 F 3.2  
4:00 195.0 F 8.0  
6:00 213.4 F 13.2  
8:00 223.7 F 8.0  
10:00 238.4 F 13.3  
12:00 238.4 F 10.9  
14:00 236.9 F 9.2  
16:00 251.3 F 16.2  
18:00 263.4 F 23.2  
20:00 271.8 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.6 F 29.3  
2:00 271.7 F 28.5  
3:00 272.2 F 29.3  
4:00 272.3 F 29.0  
  
Min 271.4 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.5 F 0.0  
5:00 220.1 F 0.0  
10:00 238.1 F 0.0  
15:00 233.8 F 0.0  
20:00 227.0 F 0.0  
25:00 223.4 F 0.0  
30:00 222.3 F 0.0  
Drying Complete

FILLING: 1:45  
HEATING: 20:06  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:05  
CYCLE COMPLETE

02/15/2023 10:56 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_