
Midmark M11 - v1.0.0
Total Cycles: 4252

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2024 11:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.7 F 0.0
2:00 155.2 F 2.5
4:00 165.8 F 5.6
6:00 194.7 F 10.0
8:00 219.6 F 5.7
10:00 230.0 F 8.0
12:00 233.2 F 12.7
14:00 236.8 F 9.4
16:00 238.7 F 9.6
18:00 246.9 F 13.6
20:00 258.4 F 19.8
22:00 268.8 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 271.9 F 28.5
2:00 272.4 F 28.7
3:00 272.6 F 28.9
4:00 272.5 F 29.2

Min 271.3 F 28.5
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 226.9 F 0.0
10:00 215.8 F 0.0
15:00 225.3 F 0.0
20:00 198.5 F 0.0
25:00 181.4 F 0.0
30:00 179.1 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 23:50
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:34
CYCLE COMPLETE

01/31/2024 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____