
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001523
10:21 23/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:48
STERILIZING 3:55
136.0 C 220kPa 3:55
136.2 C 222kPa 4:25
136.4 C 224kPa 4:55
136.3 C 225kPa 5:25
136.3 C 225kPa 5:55
136.4 C 225kPa 6:25
136.3 C 225kPa 6:55
136.3 C 225kPa 7:25
136.3 C 225kPa 7:25
VENTING 7:25
AIR DRYING 8:19
CYCLE COMPLETE 8:37

Digital Signature #
1E4994356EDCDE97
