
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002497
9:59 10/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:36
STERILIZING 3:48
136.0 C 220kPa 3:48
136.1 C 224kPa 4:18
136.3 C 225kPa 4:48
136.4 C 225kPa 5:18
136.4 C 226kPa 5:48
136.4 C 226kPa 6:18
136.4 C 226kPa 6:48
136.4 C 226kPa 7:18
136.4 C 226kPa 7:18
VENTING 7:18
AIR DRYING 8:14
CYCLE COMPLETE 36:25

Digital Signature #
193A3C0CDD76ABA2
