
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000278
14:16 01/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 3:10
STERILIZING 4:38
136.0 C 227kPa 4:38
136.4 C 233kPa 5:08
136.4 C 233kPa 5:38
136.4 C 233kPa 6:08
136.3 C 233kPa 6:38
136.2 C 233kPa 7:08
136.1 C 231kPa 7:38
136.4 C 233kPa 8:08
136.4 C 233kPa 8:08
VENTING 8:08
AIR DRYING 9:17
CYCLE COMPLETE 20:28

Digital Signature #
5F1DBA578507C2BD
