

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4138

Name: RITTER 1  
Sterilizer ID: V2178198  
12/20/2023 10:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.6 F 0.0  
2:00 151.4 F 1.5  
4:00 161.3 F 4.2  
6:00 195.3 F 8.0  
8:00 206.5 F 12.6  
10:00 218.5 F 5.6  
12:00 227.9 F 10.0  
14:00 236.2 F 8.9  
16:00 242.4 F 11.5  
18:00 238.6 F 9.5  
20:00 251.3 F 15.7  
22:00 262.4 F 21.9  
24:00 272.4 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 271.9 F 28.5  
2:00 272.5 F 28.9  
3:00 271.6 F 28.7  
4:00 271.8 F 28.5

Min 271.2 F 28.3  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.3 F 0.0  
5:00 193.9 F 0.0  
10:00 224.1 F 0.0  
15:00 232.8 F 0.0  
20:00 219.4 F 0.0  
25:00 229.0 F 0.0  
30:00 205.2 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 25:11  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:04:03  
CYCLE COMPLETE

12/20/2023 11:40 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_