
Midmark M11 - v1.0.0
Total Cycles: 2965

Name: RITTER 1
Sterilizer ID: V2178198
03/02/2023 08:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	79.3 F	0.0
2:00	79.8 F	0.0
4:00	88.1 F	0.0
6:00	122.2 F	0.2
8:00	152.1 F	1.9
10:00	181.9 F	4.2
12:00	199.7 F	7.2

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 209F
CHAMBER PRESS 10.3PSI
** ITEMS NOT STERILE **

03/02/2023 08:46 AM