
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006739
9:44 14/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:21
STERILIZING 4:12
136.0 C 220kPa 4:12
135.8 C 221kPa 4:42
136.3 C 223kPa 5:12
136.4 C 223kPa 5:42
136.1 C 217kPa 6:12
136.0 C 217kPa 6:42
135.9 C 218kPa 7:12
136.4 C 221kPa 7:42
136.4 C 221kPa 7:42
VENTING 7:43
AIR DRYING 8:34
CYCLE COMPLETE 11:55

Digital Signature #
F85374BBF16CE2C8
