
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000054
10:50 28/09/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:34
PRESSURIZING 7:59
STERILIZING 9:59
136.0 C 223kPa 9:59
136.4 C 227kPa 10:29
136.4 C 226kPa 10:59
136.4 C 226kPa 11:29
136.4 C 226kPa 12:00
136.5 C 227kPa 12:30
136.2 C 226kPa 13:00
136.3 C 226kPa 13:29
VENTING 13:30
AIR DRYING 14:45
CYCLE COMPLETE 44:45

Digital Signature #
6BF9A6D17AEBEBF3
