
Midmark M11 - v1.0.5
Total Cycles: 85

Name: RITTER2
Sterilizer ID: V2313228
06/09/2021 02:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.4 F 0.0
2:00 136.7 F 2.0
4:00 151.6 F 5.2
6:00 179.2 F 9.5
8:00 197.0 F 7.0
10:00 216.4 F 7.8
12:00 230.4 F 12.9
14:00 233.3 F 10.3
16:00 236.3 F 8.8
18:00 248.1 F 15.3
20:00 260.1 F 21.8
22:00 270.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.6
1:00 272.4 F 30.0
2:00 271.7 F 29.7
3:00 272.3 F 30.0
4:00 271.6 F 29.5

Min 271.5 F 29.4
Max 272.5 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.8 F 0.0
5:00 212.7 F 0.0
10:00 222.4 F 0.0
15:00 219.5 F 0.0
20:00 214.1 F 0.0
25:00 214.8 F 0.0
30:00 212.2 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:49
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:16
CYCLE COMPLETE

06/09/2021 03:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____