
Midmark M11 - v1.0.0
Total Cycles: 817

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2021 03:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 217.5 F 0.0
2:00 201.7 F 3.6
4:00 201.2 F 8.5
6:00 221.0 F 9.7
8:00 229.4 F 9.9
10:00 233.4 F 8.3
12:00 244.7 F 12.7
14:00 244.7 F 12.7
16:00 257.9 F 19.6
18:00 269.4 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.4 F 29.2
2:00 271.6 F 28.5
3:00 272.5 F 29.3
4:00 272.1 F 28.7

Min 271.4 F 28.5
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 233.6 F 0.0
10:00 228.3 F 0.0
15:00 230.2 F 0.0
20:00 190.7 F 0.0
25:00 185.7 F 0.0
30:00 177.3 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:07
CYCLE COMPLETE

04/27/2021 04:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____