
Midmark M11 - v1.0.5
Total Cycles: 161

Name: RITTER2
Sterilizer ID: V2313228
07/05/2021 12:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.9 F 0.0
2:00 196.2 F 3.8
4:00 203.2 F 9.3
6:00 222.7 F 8.5
8:00 231.3 F 10.8
10:00 233.4 F 8.7
12:00 239.3 F 10.5
14:00 248.1 F 14.7
16:00 260.3 F 21.6
18:00 271.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.4 F 29.6
2:00 271.3 F 28.9
3:00 272.4 F 29.6
4:00 271.6 F 28.9

Min 271.2 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 224.1 F 0.0
10:00 229.4 F 0.0
15:00 222.3 F 0.0
20:00 215.9 F 0.0
25:00 209.7 F 0.0
30:00 205.1 F 0.0
Drying Complete

FILLING: 2:29
HEATING: 18:39
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:19
CYCLE COMPLETE

07/05/2021 01:32 PM
ACCEPTED: X
REJECTED:
Operator: U01