
Midmark M11 - v1.0.5
Total Cycles: 741

Name: RITTER2
Sterilizer ID: V2313228
02/01/2022 07:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.3 F 0.0
2:00 77.8 F 0.0
4:00 88.2 F 0.0
6:00 119.6 F 0.2
8:00 152.2 F 2.5
10:00 178.3 F 5.7
12:00 196.7 F 9.4
14:00 212.4 F 11.0
16:00 223.5 F 7.3
18:00 234.5 F 11.8
20:00 234.5 F 8.9
22:00 240.5 F 10.5
24:00 246.7 F 13.8
26:00 258.9 F 20.4
28:00 269.8 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.2 F 29.1
2:00 271.9 F 29.3
3:00 271.9 F 28.8
4:00 272.1 F 29.5

Min 271.4 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 225.6 F 0.0
10:00 198.9 F 0.0
15:00 196.7 F 0.0
20:00 194.7 F 0.0
25:00 216.7 F 0.0
30:00 224.8 F 0.0

Drying Complete

FILLING: 1:07
HEATING: 28:54
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:07:04
CYCLE COMPLETE

02/01/2022 08:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____