

-----  
Midmark M11 - v1.0.5  
Total Cycles: 110

Name: RITTER2  
Sterilizer ID: V2313228  
06/17/2021 07:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.9 F 0.0  
2:00 78.3 F 0.1  
4:00 83.0 F 0.8  
6:00 96.6 F 2.6  
8:00 123.1 F 5.3  
10:00 153.9 F 8.6  
12:00 183.0 F 12.1  
14:00 198.0 F 3.7  
16:00 211.1 F 7.0  
18:00 222.9 F 10.7  
20:00 226.6 F 8.6  
22:00 236.8 F 11.0  
24:00 233.6 F 8.3  
26:00 247.0 F 14.1  
28:00 257.6 F 19.9  
30:00 267.7 F 26.2  
32:00 273.0 F 30.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.4  
1:00 272.8 F 30.2  
2:00 272.1 F 29.6  
3:00 271.3 F 29.5  
4:00 272.7 F 30.2  
  
Min 271.1 F 29.2  
Max 273.5 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.6 F 0.0  
5:00 189.1 F 0.0  
10:00 204.4 F 0.0  
15:00 211.2 F 0.0  
20:00 217.1 F 0.0  
25:00 197.3 F 0.0  
30:00 196.3 F 0.0  
  
Drying Complete

FILLING: 0:59  
HEATING: 32:26  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:10:40  
CYCLE COMPLETE

06/17/2021 08:58 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_