
Midmark M11 - v1.0.0
Total Cycles: 3542

Name: RITTER 1
Sterilizer ID: V2178198
07/18/2023 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.8 F 0.0
2:00 195.4 F 2.6
4:00 197.5 F 6.6
6:00 212.7 F 11.5
8:00 222.5 F 6.2
10:00 236.0 F 11.2
12:00 234.6 F 7.7
14:00 244.8 F 13.6
16:00 243.0 F 11.8
18:00 255.0 F 18.0
20:00 265.4 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 29.1
2:00 272.0 F 28.5
3:00 271.7 F 28.8
4:00 272.5 F 29.0

Min 271.4 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.5 F 0.0
5:00 221.9 F 0.0
10:00 233.0 F 0.0
15:00 225.4 F 0.0
20:00 224.5 F 0.0
25:00 223.5 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 21:55
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:50
CYCLE COMPLETE

07/18/2023 11:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____