
Midmark M11 - v1.0.0
Total Cycles: 3818

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2023 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.5 F 0.0
2:00 174.0 F 2.6
4:00 198.0 F 7.0
6:00 218.3 F 12.3
8:00 226.6 F 7.2
10:00 236.0 F 12.4
12:00 236.8 F 9.7
14:00 238.9 F 8.9
16:00 248.9 F 14.5
18:00 260.9 F 21.1
20:00 271.5 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.1 F 28.4
2:00 271.8 F 28.7
3:00 271.7 F 28.2
4:00 272.1 F 28.8

Min 271.0 F 28.0
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 200.3 F 0.0
10:00 207.1 F 0.0
15:00 213.5 F 0.0
20:00 214.8 F 0.0
25:00 216.8 F 0.0
30:00 212.0 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

09/26/2023 11:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____