
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005797
10:19 20/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:11
STERILIZING 4:19
136.4 C 225kPa 4:19
136.4 C 225kPa 4:49
136.0 C 222kPa 5:19
136.2 C 222kPa 5:49
136.2 C 222kPa 6:19
136.2 C 220kPa 6:49
136.4 C 225kPa 7:19
136.4 C 225kPa 7:49
136.4 C 225kPa 7:49
VENTING 7:49
AIR DRYING 8:40
CYCLE COMPLETE 8:46

Digital Signature #
A171E1353B66A800
