
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 002691
13:54 15/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 3:00
STERILIZING 4:04
136.0 C 220kPa 4:04
136.2 C 223kPa 4:34
136.0 C 217kPa 5:04
136.3 C 223kPa 5:34
136.4 C 225kPa 6:04
136.4 C 225kPa 6:34
136.4 C 223kPa 7:04
136.3 C 223kPa 7:34
136.3 C 223kPa 7:34
VENTING 7:34
AIR DRYING 8:31
CYCLE COMPLETE 17:43

Digital Signature #
4BE90FEE1ACF33AA
