
Midmark M11 - v1.0.5
Total Cycles: 3708

Name: RITTER2
Sterilizer ID: V2313228
06/17/2024 05:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 90.0 F 0.0
2:00 90.3 F 0.3
4:00 109.3 F 1.9
6:00 143.6 F 4.6
8:00 166.6 F 7.7
10:00 188.6 F 10.8
12:00 205.3 F 5.6
14:00 216.3 F 8.0
16:00 224.6 F 12.2
18:00 229.1 F 7.9
20:00 241.0 F 12.9
22:00 239.8 F 11.0
24:00 251.5 F 16.8
26:00 261.8 F 22.9
28:00 271.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.6 F 30.2
2:00 272.1 F 29.6
3:00 271.3 F 29.4
4:00 272.5 F 30.2

Min 271.2 F 29.2
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 207.2 F 0.0
10:00 209.1 F 0.0
15:00 211.1 F 0.0
20:00 206.5 F 0.0
25:00 222.3 F 0.0
30:00 234.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 28:39
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:07:14
CYCLE COMPLETE

06/17/2024 06:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____