
Midmark M11 - v1.0.5
Total Cycles: 4234

Name: RITTER2
Sterilizer ID: V2313228
12/20/2024 02:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.8 F 0.0
2:00 106.6 F 0.1
4:00 114.6 F 1.7
6:00 149.3 F 4.5
8:00 167.7 F 7.8
10:00 186.5 F 11.5
12:00 206.9 F 4.3
14:00 219.0 F 8.5
16:00 230.4 F 12.7
18:00 231.6 F 9.3
20:00 237.8 F 10.7
22:00 243.3 F 12.8
24:00 254.4 F 18.5
26:00 264.3 F 24.6
28:00 272.1 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.6
1:00 272.9 F 30.2
2:00 271.8 F 29.5
3:00 272.5 F 30.4
4:00 272.5 F 29.9

Min 271.5 F 29.5
Max 273.1 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 196.0 F 0.0
10:00 187.2 F 0.0
15:00 182.2 F 0.0
20:00 179.6 F 0.0
25:00 179.0 F 0.0
30:00 177.7 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 28:13
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:07:04
CYCLE COMPLETE

12/20/2024 03:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____