
Midmark M11 - v1.0.0
Total Cycles: 2462

Name: RITTER 1
Sterilizer ID: V2178198
10/20/2022 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.7 F 0.0
2:00 144.0 F 1.4
4:00 178.5 F 4.7
6:00 199.6 F 8.9
8:00 221.6 F 13.5
10:00 226.1 F 6.9
12:00 237.2 F 11.6
14:00 231.2 F 7.7
16:00 245.2 F 13.4
18:00 242.0 F 11.5
20:00 254.6 F 17.9
22:00 265.8 F 24.7
24:00 272.2 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 272.6 F 29.2
2:00 272.6 F 29.5
3:00 271.9 F 29.6
4:00 272.2 F 29.1

Min 271.7 F 29.0
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 183.3 F 0.0
10:00 237.3 F 0.0
15:00 244.9 F 0.0
20:00 228.6 F 0.0
25:00 221.9 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:27
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:02:44
CYCLE COMPLETE

10/20/2022 10:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____