
Midmark M11 - v1.0.0
Total Cycles: 2781

Name: RITTER 1
Sterilizer ID: V2178198
01/13/2023 01:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.2 F 0.0
2:00 194.9 F 2.1
4:00 196.6 F 5.9
6:00 216.2 F 10.9
8:00 222.7 F 5.3
10:00 231.1 F 10.6
12:00 232.2 F 7.4
14:00 245.4 F 12.8
16:00 244.1 F 12.4
18:00 256.3 F 18.8
20:00 267.4 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.3 F 29.3
2:00 272.2 F 29.0
3:00 271.5 F 28.4
4:00 271.3 F 28.6

Min 271.1 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 237.6 F 0.0
10:00 230.4 F 0.0
15:00 223.7 F 0.0
20:00 225.0 F 0.0
25:00 226.6 F 0.0
30:00 225.6 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 21:31
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:22
CYCLE COMPLETE

01/13/2023 02:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____