

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2671

Name: RITTER2  
Sterilizer ID: V2313228  
07/19/2023 03:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.1 F 0.0  
2:00 190.0 F 3.5  
4:00 202.5 F 8.4  
6:00 217.7 F 11.8  
8:00 228.5 F 9.0  
10:00 236.1 F 10.2  
12:00 242.1 F 12.6  
14:00 242.2 F 11.8  
16:00 255.1 F 18.6  
18:00 266.6 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.5 F 29.6  
2:00 271.4 F 29.2  
3:00 272.5 F 29.7  
4:00 271.6 F 29.0  
  
Min 271.4 F 29.0  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.0 F 0.0  
5:00 221.9 F 0.0  
10:00 214.0 F 0.0  
15:00 217.1 F 0.0  
20:00 220.9 F 0.0  
25:00 227.7 F 0.0  
30:00 226.8 F 0.0  
Drying Complete

FILLING: 2:10  
HEATING: 19:33  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:60  
CYCLE COMPLETE

07/19/2023 04:07 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_