
Midmark M11 - v1.0.0
Total Cycles: 898

Name: RITTER 1
Sterilizer ID: V2178198
06/15/2021 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.5 F 0.0
2:00 144.8 F 1.4
4:00 164.1 F 3.9
6:00 186.1 F 7.1
8:00 203.4 F 11.3
10:00 213.9 F 4.4
12:00 227.4 F 9.0
14:00 237.5 F 11.9
16:00 239.0 F 10.8
18:00 236.1 F 8.8
20:00 249.6 F 15.1
22:00 260.9 F 21.4
24:00 271.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.4
1:00 272.2 F 29.0
2:00 272.4 F 28.8
3:00 271.6 F 28.3
4:00 272.5 F 29.2

Min 271.6 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.9 F 0.0
5:00 212.9 F 0.0
10:00 200.9 F 0.0
15:00 201.7 F 0.0
20:00 211.4 F 0.0
25:00 232.0 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:34
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:03:11
CYCLE COMPLETE

06/15/2021 11:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____