

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4881

Name: RITTER 1  
Sterilizer ID: V2178198  
09/09/2024 03:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 125.6 F 0.0  
2:00 129.2 F 0.1  
4:00 152.1 F 0.4  
6:00 179.0 F 1.0  
8:00 197.6 F 3.9  
10:00 211.4 F 7.3  
12:00 221.1 F 11.1  
14:00 225.9 F 6.0  
16:00 235.8 F 10.6  
18:00 234.8 F 8.2  
20:00 245.7 F 13.5  
22:00 244.3 F 12.3  
24:00 256.0 F 18.4  
26:00 265.9 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.0 F 28.6  
2:00 272.6 F 29.1  
3:00 272.6 F 28.9  
4:00 272.4 F 28.7

Min 271.5 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.1 F 0.0  
5:00 196.2 F 0.0  
10:00 197.5 F 0.0  
15:00 197.0 F 0.0  
20:00 200.6 F 0.0  
25:00 214.1 F 0.0  
30:00 215.2 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 27:59  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:06:44  
CYCLE COMPLETE

09/09/2024 04:37 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_