
Midmark M11 - v1.0.0
Total Cycles: 5260

Name: RITTER 1
Sterilizer ID: V2178198
01/22/2025 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.5 F 0.0
2:00 164.5 F 3.2
4:00 180.0 F 7.2
6:00 199.3 F 11.8
8:00 211.7 F 4.4
10:00 225.9 F 9.0
12:00 234.8 F 11.6
14:00 237.8 F 10.6
16:00 235.5 F 8.0
18:00 249.1 F 14.8
20:00 260.4 F 21.1
22:00 270.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 29.1
2:00 272.3 F 28.6
3:00 271.3 F 28.3
4:00 272.1 F 29.1

Min 271.3 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 202.9 F 0.0
10:00 191.6 F 0.0
15:00 197.1 F 0.0
20:00 200.3 F 0.0
25:00 204.6 F 0.0
30:00 207.8 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 22:48
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:02:20
CYCLE COMPLETE

01/22/2025 02:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____