
Midmark M11 - v1.0.0
Total Cycles: 3387

Name: RITTER 1
Sterilizer ID: V2178198
06/14/2023 10:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.2 F 0.0
2:00 186.6 F 1.3
4:00 207.7 F 6.7
6:00 222.1 F 11.9
8:00 229.2 F 8.0
10:00 242.0 F 13.6
12:00 241.5 F 11.8
14:00 240.2 F 10.6
16:00 253.8 F 17.4
18:00 265.4 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.3 F 28.9
2:00 271.5 F 28.7
3:00 272.3 F 28.9
4:00 271.4 F 28.5

Min 271.3 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 211.4 F 0.0
10:00 217.2 F 0.0
15:00 219.4 F 0.0
20:00 219.5 F 0.0
25:00 224.4 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 19:44
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

06/14/2023 11:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____