

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1764

Name: RITTER 1  
Sterilizer ID: V2178198  
03/29/2022 04:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.4 F 0.0  
2:00 133.4 F 0.1  
4:00 159.8 F 2.2  
6:00 184.5 F 5.4  
8:00 203.3 F 10.1  
10:00 220.1 F 6.0  
12:00 230.8 F 9.3  
14:00 236.9 F 10.0  
16:00 242.1 F 12.0  
18:00 240.2 F 10.4  
20:00 253.9 F 17.3  
22:00 265.8 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.4 F 28.8  
2:00 271.5 F 28.1  
3:00 272.5 F 29.0  
4:00 271.8 F 28.3

Min 271.4 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
5:00 211.6 F 0.0  
10:00 232.4 F 0.0  
15:00 230.3 F 0.0  
20:00 225.1 F 0.0  
25:00 225.0 F 0.0  
30:00 229.3 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 23:40  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:01:47  
CYCLE COMPLETE

03/29/2022 05:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_