
Midmark M11 - v1.0.5
Total Cycles: 216

Name: RITTER2
Sterilizer ID: V2313228
07/15/2021 12:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.3 F 0.0
2:00 166.6 F 3.7
4:00 191.5 F 8.2
6:00 211.5 F 13.6
8:00 223.0 F 8.4
10:00 236.8 F 11.5
12:00 241.2 F 12.0
14:00 240.7 F 11.1
16:00 254.0 F 17.8
18:00 265.3 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.7 F 29.3
2:00 271.5 F 29.4
3:00 272.1 F 29.1
4:00 271.5 F 29.4

Min 271.1 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.6 F 0.0
5:00 233.0 F 0.0
10:00 219.6 F 0.0
15:00 199.9 F 0.0
20:00 223.1 F 0.0
25:00 220.8 F 0.0
30:00 215.4 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 19:44
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:21
CYCLE COMPLETE

07/15/2021 12:59 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____