
Midmark M11 - v1.0.5
Total Cycles: 3153

Name: RITTER2
Sterilizer ID: V2313228
11/22/2023 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.2 F 0.0
2:00 72.3 F 0.0
4:00 100.7 F 0.3
6:00 119.5 F 1.5
8:00 146.3 F 3.5
10:00 172.2 F 6.6
12:00 188.5 F 9.8
14:00 210.3 F 11.8
16:00 218.8 F 5.6
18:00 228.2 F 9.2
20:00 236.8 F 13.4
22:00 235.4 F 9.9
24:00 238.1 F 9.5
26:00 246.6 F 14.1
28:00 257.9 F 20.3
30:00 268.4 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.7 F 29.7
2:00 271.6 F 29.2
3:00 272.8 F 29.7
4:00 271.5 F 29.0

Min 271.5 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 198.7 F 0.0
10:00 194.3 F 0.0
15:00 190.5 F 0.0
20:00 193.7 F 0.0
25:00 199.4 F 0.0
30:00 200.5 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:13
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:09:28
CYCLE COMPLETE

11/22/2023 11:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____