

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.1uS / 0.7ppm  
CYCLE NUMBER 001554  
15:03 11/03/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:24  
PRESSURIZING 6:07  
STERILIZING 7:41  
136.0 C 220kPa 7:41  
136.1 C 220kPa 8:12  
136.1 C 220kPa 8:42  
136.4 C 221kPa 9:12  
136.4 C 220kPa 9:42  
135.9 C 217kPa 10:12  
136.1 C 221kPa 10:42  
136.4 C 222kPa 11:12  
VENTING 11:12  
AIR DRYING 12:21  
CYCLE COMPLETE 39:15

-----  
Digital Signature #  
576E73F995A129D2  
-----