

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1720

Name: RITTER 1  
Sterilizer ID: V2178198  
03/16/2022 07:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.8 F 0.0  
2:00 76.0 F 0.0  
4:00 106.1 F 0.1  
6:00 135.8 F 0.9  
8:00 158.2 F 2.9  
10:00 182.5 F 5.6  
12:00 199.0 F 9.0  
14:00 212.5 F 12.7  
16:00 220.4 F 5.0  
18:00 231.1 F 9.2  
20:00 239.9 F 11.4  
22:00 240.5 F 10.8  
24:00 236.3 F 8.5  
26:00 251.0 F 15.4  
28:00 262.3 F 21.9  
30:00 272.7 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 271.5 F 27.7  
2:00 271.9 F 27.9  
3:00 272.3 F 28.3  
4:00 272.7 F 28.6

Min 271.4 F 27.7  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.1 F 0.0  
5:00 203.1 F 0.0  
10:00 206.9 F 0.0  
15:00 200.6 F 0.0  
20:00 192.2 F 0.0  
25:00 209.8 F 0.0  
30:00 215.0 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 30:18  
STERILIZING: 4:00  
VENTING: 2:48  
DRYING: 30:00  
TOTAL CYCLE: 01:07:58  
CYCLE COMPLETE

03/16/2022 08:09 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_