
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001380
13:04 13/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:55
PRESSURIZING 3:57
STERILIZING 5:18
136.4 C 224kPa 5:18
136.3 C 224kPa 5:48
136.1 C 221kPa 6:18
136.3 C 224kPa 6:48
136.4 C 224kPa 7:18
136.3 C 224kPa 7:48
136.4 C 222kPa 8:18
136.3 C 224kPa 8:48
136.3 C 224kPa 8:48
VENTING 8:48
AIR DRYING 9:43
CYCLE COMPLETE 16:46

Digital Signature #
32D2758E15B2EC2E
