
Midmark M11 - v1.0.0
Total Cycles: 3144

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2023 11:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.7 F 0.0
2:00 190.5 F 0.0
4:00 202.5 F 1.2
6:00 218.0 F 5.6
8:00 232.7 F 10.2
10:00 235.1 F 8.3
12:00 242.4 F 12.4
14:00 241.2 F 11.0
16:00 239.8 F 10.1
18:00 254.0 F 17.1
20:00 265.8 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.6 F 28.6
2:00 271.5 F 27.9
3:00 272.6 F 28.8
4:00 272.1 F 28.3

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 211.4 F 0.0
10:00 224.5 F 0.0
15:00 234.2 F 0.0
20:00 226.3 F 0.0
25:00 222.4 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 21:40
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

04/21/2023 12:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____