
Midmark M11 - v1.0.5
Total Cycles: 998

Name: RITTER2
Sterilizer ID: V2313228
04/18/2022 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.3 F 0.0
2:00 191.3 F 1.0
4:00 210.9 F 6.4
6:00 222.1 F 12.0
8:00 230.5 F 8.4
10:00 241.2 F 11.1
12:00 243.8 F 13.1
14:00 244.7 F 12.8
16:00 258.2 F 20.1
18:00 270.0 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.5 F 29.2
2:00 272.0 F 29.3
3:00 272.3 F 29.0
4:00 272.2 F 29.5

Min 271.6 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 208.2 F 0.0
10:00 212.2 F 0.0
15:00 217.9 F 0.0
20:00 222.6 F 0.0
25:00 229.0 F 0.0
30:00 236.0 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 18:52
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:44
CYCLE COMPLETE

04/18/2022 01:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____