

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4102

Name: RITTER2  
Sterilizer ID: V2313228  
11/08/2024 12:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 110.1 F 0.0  
2:00 109.6 F 0.0  
4:00 114.1 F 0.0  
6:00 146.7 F 0.2  
8:00 173.1 F 3.3  
10:00 202.7 F 6.6  
12:00 213.0 F 10.7  
14:00 220.0 F 6.9  
16:00 229.0 F 9.9  
18:00 232.6 F 9.4  
20:00 241.4 F 12.5  
22:00 240.8 F 11.4  
24:00 252.9 F 17.4  
26:00 263.6 F 23.8  
28:00 272.3 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.1  
1:00 272.2 F 30.1  
2:00 272.6 F 29.8  
3:00 272.4 F 30.4  
4:00 272.4 F 29.8

Min 270.9 F 28.9  
Max 273.2 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.0 F 0.0  
5:00 217.1 F 0.0  
10:00 215.5 F 0.0  
15:00 208.7 F 0.0  
20:00 213.7 F 0.0  
25:00 220.6 F 0.0  
30:00 223.4 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 28:15  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:06:52  
CYCLE COMPLETE

11/08/2024 01:47 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_