
Midmark M11 - v1.0.0
Total Cycles: 2860

Name: RITTER 1
Sterilizer ID: V2178198
02/06/2023 12:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.9 F 0.0
2:00 194.8 F 4.2
4:00 200.7 F 8.9
6:00 217.4 F 9.6
8:00 224.8 F 8.1
10:00 237.6 F 13.3
12:00 238.3 F 10.5
14:00 236.4 F 8.8
16:00 251.5 F 16.1
18:00 263.8 F 23.2
20:00 271.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.5 F 28.9
2:00 271.8 F 28.6
3:00 272.4 F 28.9
4:00 271.6 F 28.5

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 214.8 F 0.0
10:00 233.8 F 0.0
15:00 231.1 F 0.0
20:00 224.2 F 0.0
25:00 222.7 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 20:00
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

02/06/2023 01:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____